

IOP correction for Central Corneal Thickness

Central Corneal Thickness (μm)	Correction Value (mmHg)
410	+10
415	+10
420	+9
425	+9
430	+8
435	+8
440	+7
445	+7
455	+6
465	+6
475	+5
485	+4
495	+4
505	+3
515	+2
525	+1
535	+1
545	0
555	-1
565	-1
575	-2
585	-3
595	-4
605	-4
615	-5
625	-6
635	-6
645	-7

References

Ehlers N., Bramsen T., Sperling S. Applanation tonometry and central corneal thickness. *Acta Ophthalmol.* 1975;53:34-43

[glaucoma, cornea](#)

From:

<https://childreneye.org/wiki/> - **Children's Eye Wiki**

Permanent link:

https://childreneye.org/wiki/doku.php?id=iop_correction_corneal_thickness

Last update: **2020/11/22 20:44**

